



Ref. Certif. No.

SG PSB-MD-00319

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product

Switching power supply unit  
(MEDICAL POWER SUPPLY)

Name and address of the applicant

**Shenzhen MEGMEET Electrical Co., Ltd.**  
5th Floor, Tower A, B,C501-C503,D,E  
Unis Inforport, No.13 Langshan Road  
North Section, Hi-Tech Industrial Park  
Nanshan District  
518057 Shenzhen  
PEOPLE'S REPUBLIC OF CHINA

Name and address of the manufacturer

Shenzhen MEGMEET Electrical Co., Ltd.  
5th Floor, Tower A, B,C501-C503,D,E, Unis Inforport, No.13  
Langshan Road, North Section, Hi-Tech Industrial Park, Nanshan  
District, 518057 Shenzhen, PEOPLE'S REPUBLIC OF CHINA

Name and address of the factory

ZHUZHOU MEGMEET ELECTRICAL CO., LTD  
1728# TAISHAN RD., TIANYUAN DISTRICT, 412007 ZHUZHOU,  
HU'NAN, PEOPLE'S REPUBLIC OF CHINA

Ratings and principal characteristics

Rated Input: 100-240VAC, 50/60Hz, 2.2-1.0A  
Rated DC output: See test report for details  
Protection Class: I  
Degree of Protection: IPX0

Trade mark

MEGMEET

Model/type Ref.

MKP150-xxA-abc-zzzz  
(See test report for details of model description)

Additional information (if necessary)

See Test Report for National Differences and Group Differences.  
The risk management requirements of the standard were not  
addressed.

A sample of the product was tested and found  
to be in conformity with

IEC 60601-1:2005  
IEC 60601-1:2005/AMD1:2012

as shown in the Test Report Ref. No.  
which forms part of this certificate

211-27190021-000

This CB Test Certificate is issued by the National Certification Body

CBS 075834 0202 Rev. 00

Date, 2019-09-04

( Yager Bi )

Page 1 of 1

TÜV SÜD PSB Pte Ltd • 1 Science Park Drive • Singapore 118221



PSB Singapore