



Ref. Certif. No.

SG PSB-MD-00270

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Power supply
(MEDICAL AC/DC ADAPTER)

Name and address of the applicant

Shenzhen MEGMEET Electrical Co., Ltd.
5th Floor, Tower A, B,C501-C503,D,E
Unis Inforport, No.13 Langshan Road
North Section, Hi-Tech Industrial Park
Nanshan District
518057 Shenzhen
PEOPLE'S REPUBLIC OF CHINA

Name and address of the manufacturer

Shenzhen MEGMEET Electrical Co., Ltd.
5th Floor, Tower A, B,C501-C503,D,E, Unis Inforport, No.13
Langshan Road, North Section, Hi-Tech Industrial Park, Nanshan
District, 518057 Shenzhen, PEOPLE'S REPUBLIC OF CHINA

Name and address of the factory

ZHUZHOU MEGMEET ELECTRICAL CO., LTD
1728# TAISHAN RD., TIANYUAN DISTRICT, 412007 ZHUZHOU,
HU'NAN, PEOPLE'S REPUBLIC OF CHINA

Ratings and principal characteristics

Rated input: 100-240VAC, 50/60Hz, 1.5A
Rated output: 5VDC, 3A or 9VDC, 3A or 10.4VDC, 3A or
15VDC, 3A or 20VDC, 3A; 60W MAX
Protection Class:
Class I (For models MANGO60S-USB-PDA and MANGO60S-USB-PDD)
Class II (For models MANGO60S-USB-PDB, MANGO60S-USB-PDC)
Degree of protection: IP20

Trade mark

MEGMEET

Model/type Ref.

MANGO60S-USB-PDA, MANGO60S-USB-PDB,
MANGO60S-USB-PDC, MANGO60S-USB-PDD

Additional information (if necessary)

See Test Report for National Differences and Group Differences. The
risk management requirements of the standard were not addressed.A sample of the product was tested and found
to be in conformity withIEC 60601-1:2005
IEC 60601-1:2005/AMD1:2012as shown in the Test Report Ref. No.
which forms part of this certificate

211-27190005-000

This CB Test Certificate is issued by the National Certification Body

CBS 075834 0181 Rev. 00

Date, 2019-03-26

Page 1 of 1

TÜV SÜD PSB Pte Ltd • 1 Science Park Drive • Singapore 118221

(Yager Bi)



PSB Singapore