



Ref. Certif. No.

SG PSB-MD-00570M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Adaptors (MEDICAL AC/DC ADAPTER / AC/DC ADAPTER)
Name and address of the applicant	Shenzhen MEGMEET Electrical Co., Ltd. 5th Floor, Tower A, B,C501-C503,D,E Unis Inforport, No.13 Langshan Road North Section, Hi-Tech Industrial Park Nanshan District 518057 Shenzhen PEOPLE'S REPUBLIC OF CHINA
Name and address of the manufacturer	Shenzhen MEGMEET Electrical Co., Ltd. 5th Floor, Tower A, B,C501-C503,D,E, Unis Inforport, No.13 Langshan Road, North Section, Hi-Tech Industrial Park, Nanshan District, 518057 Shenzhen, PEOPLE'S REPUBLIC OF CHINA
Name and address of the factory	ZHUZHOU MEGMEET ELECTRICAL CO., LTD. 1728# TAISHAN RD, TIANYUAN DISTRICT, 412007 ZHUZHOU, HU'NAN, PEOPLE'S REPUBLIC OF CHINA
Ratings and principal characteristics	Rated Input: 100-240VAC, 50/60Hz, 2.0-1.0A or 2.0A MAX or 2.2A MAX or 2.2-1.0A Rated Output: See test report for details Protection Class: I
Trademark / Brand (if any)	MEGMEET or N/A
Model/type Ref.	MANGO120S-xyzz-abc, MANGO120S-xxxyzz-abc, MANGO150S-xyzz-abc, MANGO150S-xxxyzz-abc, MANGO150S-xyzz, MANGO120S-xyzz, MANGO150S-xxxyzz, MANGO120S-xxxyzz (See test report for details of model description)
Additional information (if necessary)	Certificate SG PSB-MD-00570 issued on 2022-06-22 is replaced by this version due to technical changes.
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1:2005/AMD2:2020
as shown in the Test Report Ref. No. which forms part of this certificate	211-900225-100

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 075834 0312 Rev. 01

Date, 2024-02-07

(Yager Bi)

TÜV SÜD PSB Pte Ltd 15 International Business Park TÜV SÜD @ IBP Singapore 609937



PSB Singapore



Ref. Certif. No.

SG PSB-MD-00570M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

Name and address of the factory (continued)

Altatronic International Co., Ltd.
7/375 Moo 6, Tambon Mabyangporn, Pluak Daeng, Rayong 21140,
THAILAND

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 075834 0312 Rev. 01

Date, 2024-02-07

(Yager Bi)



TÜV SÜD PSB Pte Ltd 15 International Business Park TÜV SÜD @ IBP Singapore 609937

PSB Singapore