

SG PSB-MD-00455

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

**CB TEST CERTIFICATE** 

Product General Medical Devices

Power supply

Name and address of the applicant Shenzhen MEGMEET Electrical Co., Ltd.

5th Floor, Tower A, B,C501-C503,D,E Unis Inforport, No.13 Langshan Road North Section, Hi-Tech Industrial Park

Nanshan District 518057 Shenzhen

PEOPLE'S REPUBLIC OF CHINA

Name and address of the manufacturer Shenzhen MEGMEET Electrical Co., Ltd.

5th Floor, Tower A, B,C501-C503,D,E, Unis Inforport, No.13 Langshan Road, North Section, Hi-Tech Industrial Park, Nanshan District, 518057

Shenzhen, PEOPLE'S REPUBLIC OF CHINA

Name and address of the factory ZHUZHOU MEGMEET ELECTRICAL CO., LTD

1728# TAISHAN RD, TIANYUAN DISTRICT, 412007 ZHUZHOU, HU'NAN,

PEOPLE'S REPUBLIC OF CHINA

Megmeet (Thailand) Co., Ltd.

7/375 Moo6, Mabyangporn, Plueakdaeng, Rayong 21140, THAILAND

Ratings and principal characteristics AC Input: 100-240VAC, 50/60Hz, 2.0-1.0A

DC Output: +5V/3A, 15W or +9V/3A, 27W or +10.4V/3A, 31W or +12V/3A, 36W or

+15V/3A, 45W or +20V/5A, 100W

(There are six kinds of DC output for Power supply, one or more than one kind(s)

DC output can be disabled by manufacturer for marketing purpose)

Trade mark (if any) MEGMEET or N/A

Model/type Ref. Refer to the next page

Additional information (if necessary) In addition, all applicable Clauses of IEC 62366 have be addressed.

A sample of the product was tested and found IEC 60601-1:2005

to be in conformity with IEC 60601-1:2005/AMD1:2012

IEC 60601-1-6:2010

IEC 60601-1-6:2010/AMD1:2013

IEC 60601-1-11:2015

as shown in the Test Report Ref. No. which forms part of this certificate

211-27200031-000

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 075834 0249 Rev. 00 Date, 2021-03-04

Con



(Vivian Li )





**SG PSB-MD-00455** 

IEC	SYSTEM	FOR	MUTUAL	RECOGNITION	OF	TEST	CERTIFICATES	FOR	ELECTRICAL	<b>EQUIPMENT</b>
(IECI	EE) CB SC	HEME								

Models (continued)

MANGO100S-USB-abcdef

(a, b, c, d, e, f = A-Z or 0-9 or blank, stands for the customer identity)

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 075834 0249 Rev. 00

Date, 2021-03-04

(Vivian Li)



TÜV SÜD @ IBP Singapore 609937 PSB Singapore